



Certificate of Analysis

Product Name: ObaGel®

SKU: OS-301-01, OS-301-02, OS-301-03

Lot #: ER20230202

Analysis Date: 2/22/2023

Publication Date: 3/1/2023

Results

<i>Analysis</i>	<i>Specification</i>	<i>Value</i>	<i>Unit</i>	<i>Method</i>
<i>Sterility - Fungal</i>	Clear	Pass		Microscopy
<i>Sterility - Bacterial</i>	Clear	Pass		Inoculation
<i>Gelation</i>	30 minutes, 37°C	Pass		Incubation
<i>Color</i>	>0.2 @ 420nm	0.416	Optical density (OD) undiluted sample	Microplate reader assay, absorbance
<i>Protein Content</i>	>10 mg/mL	63.23 ± 1.13	mg/mL (average ± stdev)	BCA assay
<i>Biocompatibility</i>	>0% indicating metabolic activity	29.43 ± 1.95	% reduction, 24-48hr (average ± stdev)	alamarBlue proliferation assay

Conforms to quality standards

Chief Scientific Officer

A handwritten signature in blue ink, appearing to read "Cecilia Sanchez", written over a horizontal line.

Cecilia Sanchez